



Certificate of Compliance

Date	08/04/2026
Product name	VENABIO® TECHNOSIL
Product Prefix	92008544*
Material	VENA® SILICONE TECH 70
Manufacturing Location	Venair Biotech SLU: C/ Alguer, 26-28 or C/Perpinya 27 (different entrances of the same building) - 08226 Terrassa (Barcelona) - Spain

Properties	Unit	Inspection Method	Value
Hardness	Shore A	ISO 868	70 ± 5
Tensile strength	N/mm ²	ISO 37	≥7
Elongation	%	ISO 37	>250
Tear strength B	kN/m	ISO 34-1	>20
Density	g/cm ³	ISO 2781	1,19 ± 0,03

Certified information

Process	Quality management system certified ISO 9001:2015, ISO 14001:2015, ISO 13485:2016. Produced in ISO7 classified clean room according to ISO14644 in Venair's Spanish manufacturing plant.
----------------	--

Product	FDA Standard 21 CFR 177.2600 Section (e) and (f)
	Regulation (EU) No. 10/2011
	USP In Vivo <88> Class VI at 121°C
	ISO 10993-5 & USP In vitro <87>
	ISO 10993-4, 10, 11 & 23
	BfR XV
	EP 3.1.9

Approval

6ab6e60f-c2fc-389c-a5e0-9f15bf95526b
2026-04-08 12:54:29 UTC

Venair Biotech Quality Department

Venair Biotech SLU
C/Perpinya 27 – 08226 – Terrassa (Barcelona) - Spain

It is the user's responsibility to ensure the suitability and safety of the VENAIR products for all intended uses. All the tests must be conducted in accordance with applicable regulatory requirements to determine the safety and effectiveness for use of the hoses in any particular application. User assumes all other risk, if any, including the risk of injury, loss or damage, direct or consequential, arising out of the use, misuse, or inability to use, this product. Vena, Venair and the Venair logo are trademarks of Venair Ibérica SA.